

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072122\
 Data File : VU049821.D
 Acq On : 21 Jul 2022 11:08
 Operator : SY/MD
 Sample : N3792-08
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DBUN2

Quant Time: Jul 22 04:11:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM071122WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jul 22 03:08:56 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	173727	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	168015	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	69629	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	61499	41.956	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	83.920%
7) Chloroethane-d5	1.900	69	49741	46.946	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	93.900%
11) 1,1-Dichloroethene-d2	2.559	63	79306	30.630	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	61.260%
21) 2-Butanone-d5	4.620	46	135381	108.191	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	108.190%
24) Chloroform-d	5.060	84	128184	49.374	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.740%
26) 1,2-Dichloroethane-d4	5.700	65	90132	51.296	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.600%
32) Benzene-d6	5.726	84	234584	45.959	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.920%
36) 1,2-Dichloropropane-d6	6.690	67	86968	50.656	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	101.320%
41) Toluene-d8	7.896	98	187743	42.237	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	84.480%
43) trans-1,3-Dichloroprop...	8.179	79	31243	43.192	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	86.380%
47) 2-Hexanone-d5	8.633	63	66515	86.128	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	86.130%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	122515	51.000	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	102.000%
66) 1,2-Dichlorobenzene-d4	12.192	152	66550	50.693	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	101.380%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072122\
Data File : VU049821.D
Acq On : 21 Jul 2022 11:08
Operator : SY/MD
Sample : N3792-08
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DBUN2

Quant Time: Jul 22 04:11:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM071122WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jul 22 03:08:56 2022
Response via : Initial Calibration

