

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072122\
 Data File : VU049825.D
 Acq On : 21 Jul 2022 13:24
 Operator : SY/MD
 Sample : N3792-11
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DBUN5

Quant Time: Jul 22 04:12:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM071122WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jul 22 03:08:56 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	171926	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	165390	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	67151	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	62302	42.949	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	85.900%
7) Chloroethane-d5	1.900	69	47932	45.712	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	91.420%
11) 1,1-Dichloroethene-d2	2.562	63	79738	31.119	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	62.240%
21) 2-Butanone-d5	4.620	46	144505	116.692	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	116.690%
24) Chloroform-d	5.063	84	127649	49.683	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.360%
26) 1,2-Dichloroethane-d4	5.700	65	92950	53.454	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	106.900%
32) Benzene-d6	5.726	84	234003	46.572	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.140%
36) 1,2-Dichloropropane-d6	6.690	67	85420	50.544	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	101.080%
41) Toluene-d8	7.899	98	187038	42.746	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	85.500%
43) trans-1,3-Dichloroprop...	8.179	79	34304	48.176	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	96.360%
47) 2-Hexanone-d5	8.632	63	71321	93.816	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	93.820%
56) 1,1,2,2-Tetrachloroeth...	10.755	84	125794	53.196	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	106.400%
66) 1,2-Dichlorobenzene-d4	12.195	152	64002	50.551	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	101.100%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072122\
Data File : VU049825.D
Acq On : 21 Jul 2022 13:24
Operator : SY/MD
Sample : N3792-11
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DBUN5

Quant Time: Jul 22 04:12:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM071122WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jul 22 03:08:56 2022
Response via : Initial Calibration

