

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072122\
 Data File : VU049827.D
 Acq On : 21 Jul 2022 14:11
 Operator : SY/MD
 Sample : N3792-15
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DBUN8

Quant Time: Jul 22 04:12:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM071122WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jul 22 03:08:56 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	175480	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	166895	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	67397	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.594	65	61189	41.327	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	82.660%
7) Chloroethane-d5	1.900	69	49054	45.835	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	91.660%
11) 1,1-Dichloroethene-d2	2.562	63	78131	29.875	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	59.740%#
21) 2-Butanone-d5	4.620	46	146858	116.190	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	116.190%
24) Chloroform-d	5.063	84	128945	49.171	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.340%
26) 1,2-Dichloroethane-d4	5.700	65	92334	52.024	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	104.040%
32) Benzene-d6	5.726	84	237197	46.782	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.560%
36) 1,2-Dichloropropane-d6	6.690	67	86970	50.997	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	102.000%
41) Toluene-d8	7.899	98	185699	42.057	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	84.120%
43) trans-1,3-Dichloroprop...	8.179	79	34532	48.059	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	96.120%
47) 2-Hexanone-d5	8.632	63	74099	96.592	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	96.590%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	125423	52.561	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	105.120%
66) 1,2-Dichlorobenzene-d4	12.195	152	64177	50.504	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	101.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072122\
Data File : VU049827.D
Acq On : 21 Jul 2022 14:11
Operator : SY/MD
Sample : N3792-15
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DBUN8

Quant Time: Jul 22 04:12:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM071122WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jul 22 03:08:56 2022
Response via : Initial Calibration

