

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072123\
 Data File : VU054837.D
 Acq On : 21 Jul 2023 10:06
 Operator : MD/SY
 Sample : VU0721WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK019

Quant Time: Jul 22 02:27:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM072023.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jul 21 01:10:23 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.258	114	3695	0.500	ug/L	0.00
5) Chlorobenzene-d5	9.419	117	5425	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.808	152	1664	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.597	65	1659	0.448	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	90.000%
4) 1,2-Dichloroethane-d4	5.685	65	962	0.485	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	96.000%
7) 1,2-Dichloropropane-d6	6.687	67	1283	0.526	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	106.000%
8) Toluene-d8	7.888	98	3469	0.480	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	96.000%
10) 1,1,2,2-Tetrachloroeth...	10.746	84	957	0.508	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	102.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072123\
Data File : VU054837.D
Acq On : 21 Jul 2023 10:06
Operator : MD/SY
Sample : VU0721WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK019

Quant Time: Jul 22 02:27:17 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM072023.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jul 21 01:10:23 2023
Response via : Initial Calibration

