

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU072220\
 Data File : VU039627.D
 Acq On : 22 Jul 2020 13:40
 Operator : SY/MD
 Sample : L3346-17
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK01

Quant Time: Jul 23 02:25:43 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR072120WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Jul 23 01:28:13 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.27	114	108332	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	104670	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	47948	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	21604	4.30	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	86.00%
7) Chloroethane-d5	1.93	69	20696	3.77	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	75.40%
11) 1,1-Dichloroethene-d2	2.58	63	37358	3.31	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	66.20%
20) 2-Butanone-d5	4.67	46	113306	49.96	ug/L	0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	99.92%
24) Chloroform-d	5.09	84	54071	4.24	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.80%
26) 1,2-Dichloroethane-d4	5.72	65	36529	4.34	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.80%
32) Benzene-d6	5.75	84	105436	4.19	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.80%
36) 1,2-Dichloropropane-d6	6.70	67	36616	4.42	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	88.40%
41) Toluene-d8	7.91	98	91757	3.85	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	77.00%
43) trans-1,3-Dichloropropene-	8.19	79	13675	3.80	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	76.00%
46) 2-Hexanone-d5	8.64	63	78696	44.70	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	89.40%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	32665	4.25	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	85.00%
64) 1,2-Dichlorobenzene-d4	12.19	152	39122	4.54	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	90.80%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU072220\
Data File : VU039627.D
Acq On : 22 Jul 2020 13:40
Operator : SY/MD
Sample : L3346-17
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK01

Quant Time: Jul 23 02:25:43 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR072120WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Jul 23 01:28:13 2020
Response via : Initial Calibration

