

Quantitation Report (Qedit)




Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU072220\
Data File : VU039622.D
Acq On : 22 Jul 2020 11:42
Operator : SY/MD
Sample : L3348-13
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GAXZ5

Manual Integrations
APPROVED

MMDadoda
7/23/2020 12:09:34 PM

Quant Time: Jul 23 01:29:59 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR072120WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Jul 23 01:28:13 2020
Response via : Initial Calibration



TIC: VU039622.D

(13) Acetone (T)

2.658min (-0.003) 1.30ug/L

response 1574

Ion	Exp%	Act%
43.05	100	100
58.05	29.50	36.79
0.00	0.00	0.00
0.00	0.00	0.00

Quantitation Report (Qedit)




Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU072220\
 Data File : VU039622.D
 Acq On : 22 Jul 2020 11:42
 Operator : SY/MD
 Sample : L3348-13
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
ClientSampleId :
 GAXZ5

Manual Integrations
APPROVED

MMDadoda
 7/23/2020 12:09:34 PM

Quant Time: Jul 23 01:29:59 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR072120WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Jul 23 01:28:13 2020
 Response via : Initial Calibration



TIC: VU039622.D

(13) Acetone (T)

2.671min (+0.010) 3.45ug/L m

07/28/2020

response 4184

Ion	Exp%	Act%
43.05	100	100
58.05	29.50	13.84
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU072220\
 Data File : VU039622.D
 Acq On : 22 Jul 2020 11:42
 Operator : SY/MD
 Sample : L3348-13
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GAXZ5

Manual Integrations
 APPROVED

MMDadoda
 7/23/2020 12:09:34 PM

Quant Time: Jul 23 01:48:52 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR072120WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Jul 23 01:28:13 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.27	114	104231	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	100087	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	45536	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	33235	6.87	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	137.40%#
7) Chloroethane-d5	1.93	69	31521	5.97	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	119.40%
11) 1,1-Dichloroethene-d2	2.58	63	52216	4.80	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	96.00%
20) 2-Butanone-d5	4.66	46	120931	55.42	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	110.84%
24) Chloroform-d	5.09	84	68883	5.62	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	112.40%
26) 1,2-Dichloroethane-d4	5.72	65	42980	5.30	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	106.00%
32) Benzene-d6	5.74	84	132842	5.52	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	110.40%
36) 1,2-Dichloropropane-d6	6.71	67	43424	5.48	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	109.60%
41) Toluene-d8	7.91	98	118567	5.20	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.00%
43) trans-1,3-Dichloropropene-	8.19	79	16490	4.79	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	95.80%
46) 2-Hexanone-d5	8.64	63	84847	50.40	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	100.80%
57) 1,1,2,2-Tetrachloroethane-	10.76	84	36850	5.01	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	100.20%
64) 1,2-Dichlorobenzene-d4	12.19	152	44571	5.44	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	108.80%

Target Compounds

				Ovalue	
3) Chloromethane	1.53	50	1475	0.266 ug/L	90
13) Acetone	2.67	43	4184m	3.445 ug/L	78
16) Methylene chloride	3.06	84	2109	0.296 ug/L #	99
42) Toluene	7.98	91	2456	0.093 ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU072220\
Data File : VU039622.D
Acq On : 22 Jul 2020 11:42
Operator : SY/MD
Sample : L3348-13
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GAXZ5

Manual Integrations
APPROVED

MMDadoda
7/23/2020 12:09:34 PM

Quant Time: Jul 23 01:48:52 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR072120WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Jul 23 01:28:13 2020
Response via : Initial Calibration

