

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072221\
 Data File : VU044385.D
 Acq On : 22 Jul 2021 11:03
 Operator : SY/MD
 Sample : VU0722WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU0722WBL01

Quant Time: Jul 23 03:28:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U071621DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Fri Jul 16 14:36:39 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Fluorobenzene	6.121	96	46556	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.639	95	17355	0.924	ug/l	0.00
Spiked Amount 1.000			Recovery	=	92.000%	
68) 1,2-Dichlorobenzene-d4	12.198	152	16803	0.917	ug/l	0.00
Spiked Amount 1.000			Recovery	=	92.000%	
Target Compounds						
15) Methylene Chloride	3.054	84	6532	0.387	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072221\
Data File : VU044385.D
Acq On : 22 Jul 2021 11:03
Operator : SY/MD
Sample : VU0722WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0722WBL01

Quant Time: Jul 23 03:28:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U071621DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Fri Jul 16 14:36:39 2021
Response via : Initial Calibration

