

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072221\
 Data File : VU044384.D
 Acq On : 22 Jul 2021 10:03
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC010

Quant Time: Jul 23 03:28:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U071621DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Fri Jul 16 14:36:39 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 200%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 i	Fluorobenzene	1.000	1.000	0.0	89	0.00
2 T	Dichlorodifluoromethane	0.349	0.346	0.9	85	0.00
3 t	Chloromethane	0.380	0.375	1.3	86	0.00
4 Rt	Vinyl Chloride	0.383	0.408	-6.5	92	0.00
5 T	Bromomethane	0.204	0.208	-2.0	88	0.00
6 T	Chloroethane	0.233	0.237	-1.7	93	0.00
7 T	Trichlorofluoromethane	0.414	0.463	-11.8	95	0.00
8	1,1,2-Trichloro-1,2,2-trifl	0.262	0.291	-11.1	95	0.00
9 Rt	1,1-Dichloroethene	0.255	0.280	-9.8	94	0.00
10 t	Iodomethane	0.334	0.371	-11.1	91	0.00
11 t	Allyl Chloride	0.421	0.460	-9.3	94	0.00
12 t	Acrylonitrile	0.059	0.066	-11.9	87	0.00
13 T	Acetone	0.029	0.035	-20.7	93	0.00
14 T	Carbon Disulfide	0.845	0.894	-5.8	90	0.00
15 RT	Methylene Chloride	0.362	0.333	8.0	94	0.00
16 RT	trans-1,2-Dichloroethene	0.285	0.304	-6.7	93	0.00
17 t	1,1-Dichloroethane	0.530	0.587	-10.8	96	0.00
18 T	2-Butanone	0.063	0.076	-20.6	93	0.00
19	Cyclohexane	0.485	0.518	-6.8	92	0.00
20	Methylcyclohexane	0.506	0.534	-5.5	89	0.00
21 T	2,2-Dichloropropane	0.435	0.478	-9.9	95	0.00
22 RT	cis-1,2-Dichloroethene	0.311	0.342	-10.0	95	0.00
23 t	Diethyl Ether	0.214	0.245	-14.5	97	0.00
24 t	tert-Butyl Alcohol	0.013	0.016	-23.1	104	0.00
25 t	Methyl tert-Butyl Ether	0.713	0.796	-11.6	95	0.00
26 t	Bromochloromethane	0.132	0.145	-9.8	97	0.00
27 t	Chloroform	0.513	0.567	-10.5	98	0.00
28 RT	1,1,1-Trichloroethane	0.428	0.465	-8.6	94	0.00
29 T	1,1-Dichloropropene	0.407	0.435	-6.9	91	0.00
30 RT	Carbon Tetrachloride	0.357	0.378	-5.9	91	0.00
31 t	Isopropyl Ether	0.883	0.987	-11.8	95	0.00
32	Ethyl-t-butyl ether	0.870	0.978	-12.4	96	0.00
33	Tert-Amyl methyl ether	0.774	0.835	-7.9	94	0.00
34 t	Propionitrile	0.018	0.022	-22.2	99	0.00
35 RT	Benzene	1.182	1.258	-6.4	92	0.00
36 RT	1,2-Dichloroethane	0.339	0.380	-12.1	99	0.00
37 RT	Trichloroethene	0.297	0.311	-4.7	90	0.00
38 Rt	1,2-Dichloropropane	0.313	0.330	-5.4	91	0.00
39 t	Methacrylonitrile	0.101	0.118	-16.8	97	0.00
40 t	Methyl acrylate	0.157	0.176	-12.1	92	0.00
41 t	Tetrahydrofuran	0.048	0.051	-6.2	91	0.00
42 t	1-Chlorobutane	0.575	0.613	-6.6	93	0.00
43 T	Dibromomethane	0.155	0.170	-9.7	93	0.00
44 T	Bromodichloromethane	0.368	0.399	-8.4	93	0.00
45 T	4-Methyl-2-Pentanone	0.211	0.227	-7.6	91	0.00
46 t	t-1,4-Dichloro-2-butene	0.089	0.096	-7.9	86	0.00
47 t	Methyl methacrylate	0.172	0.182	-5.8	91	0.00
48 t	Ethyl methacrylate	0.354	0.371	-4.8	89	0.00
49 Rt	Toluene	0.744	0.791	-6.3	90	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072221\
 Data File : VU044384.D
 Acq On : 22 Jul 2021 10:03
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC010

Quant Time: Jul 23 03:28:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U071621DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Fri Jul 16 14:36:39 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 200%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
50 T	t-1,3-Dichloropropene	0.382	0.409	-7.1	88	0.00
51 T	cis-1,3-Dichloropropene	0.451	0.481	-6.7	89	0.00
52 RT	1,1,2-Trichloroethane	0.229	0.248	-8.3	94	0.00
53 t	1,3-Dichloropropane	0.420	0.449	-6.9	93	0.00
54 t	2-Hexanone	0.150	0.163	-8.7	92	0.00
55 t	Dibromochloromethane	0.248	0.269	-8.5	92	0.00
56 T	1,2-Dibromoethane	0.224	0.237	-5.8	91	0.00
57 S	4-Bromofluorobenzene	0.404	0.420	-4.0	93	0.00
58 RT	Tetrachloroethene	0.252	0.272	-7.9	92	0.00
59 Rt	Chlorobenzene	0.786	0.825	-5.0	89	0.00
60 T	1,1,1,2-Tetrachloroethane	0.262	0.280	-6.9	90	0.00
61 t	Pentachloroethane	0.209	0.211	-1.0	85	0.00
62 t	Hexachloroethane	0.219	0.233	-6.4	87	0.00
63 Rt	Ethyl Benzene	1.405	1.502	-6.9	90	0.00
64 RT	m/p-Xylenes	0.537	0.572	-6.5	90	0.00
65 RT	o-Xylene	0.518	0.558	-7.7	90	0.00
66 RT	Styrene	0.893	0.964	-8.0	90	0.00
67 t	Bromoform	0.139	0.151	-8.6	88	0.00
68 S	1,2-Dichlorobenzene-d4	0.393	0.382	2.8	86	0.00
69 T	Isopropylbenzene	1.386	1.472	-6.2	89	0.00
70 T	1,1,2,2-Tetrachloroethane	0.302	0.334	-10.6	94	0.00
71 T	1,2,3-Trichloropropane	0.259	0.250	3.5	83	0.00
72 t	Bromobenzene	0.321	0.333	-3.7	89	0.00
73 t	n-propylbenzene	0.374	0.397	-6.1	88	0.00
74 t	2-Chlorotoluene	0.321	0.330	-2.8	89	0.00
75 t	1,3,5-Trimethylbenzene	1.172	1.249	-6.6	90	0.00
76 t	4-Chlorotoluene	0.333	0.347	-4.2	90	0.00
77 t	tert-Butylbenzene	1.149	1.210	-5.3	89	0.00
78 t	1,2,4-Trimethylbenzene	1.189	1.280	-7.7	89	0.00
79 t	sec-Butylbenzene	1.564	1.680	-7.4	90	0.00
80	Nitrobenzene	0.009	0.009	0.0	75	0.00
81 t	p-Isopropyltoluene	1.295	1.386	-7.0	89	0.00
82 t	1,3-Dichlorobenzene	0.638	0.662	-3.8	89	0.00
83 Rt	1,4-Dichlorobenzene	0.644	0.667	-3.6	89	0.00
84 t	n-Butylbenzene	1.268	1.410	-11.2	91	0.00
85 Rt	1,2-Dichlorobenzene	0.618	0.649	-5.0	90	0.00
86 t	1,2-Dibromo-3-Chloropropane	0.052	0.061	-17.3	94	0.00
87 Rt	1,2,4-Trichlorobenzene	0.437	0.466	-6.6	87	0.00
88 t	Hexachlorobutadiene	0.211	0.219	-3.8	86	0.00
89 t	Naphthalene	0.874	0.971	-11.1	89	0.00
90 t	1,2,3-Trichlorobenzene	0.397	0.430	-8.3	90	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0