

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072222\
 Data File : VU049849.D
 Acq On : 22 Jul 2022 14:55
 Operator : SY/MD
 Sample : VU0722WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK095

Quant Time: Jul 23 01:05:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR072222WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Jul 23 01:04:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	88707	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	84155	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	35992	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.594	65	39101	5.445	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	108.800%
7) Chloroethane-d5	1.909	69	30549	5.083	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	101.600%
11) 1,1-Dichloroethene-d2	2.565	65	16714	5.214	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	104.200%
20) 2-Butanone-d5	4.636	46	150003	58.051	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	116.100%
24) Chloroform-d	5.063	84	80698	5.045	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.000%
26) 1,2-Dichloroethane-d4	5.703	65	49326	5.259	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.200%
32) Benzene-d6	5.729	84	147919	5.158	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.200%
36) 1,2-Dichloropropane-d6	6.690	67	52469	5.296	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	106.000%
41) Toluene-d8	7.899	98	120358	4.802	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.000%
43) trans-1,3-Dichloroprop...	8.182	79	17832	4.958	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	99.200%
46) 2-Hexanone-d5	8.636	63	83210	58.382	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	116.760%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	44812	5.386	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	107.800%
66) 1,2-Dichlorobenzene-d4	12.195	152	39300	5.290	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	105.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072222\
Data File : VU049849.D
Acq On : 22 Jul 2022 14:55
Operator : SY/MD
Sample : VU07222WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK095

Quant Time: Jul 23 01:05:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR072222WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jul 23 01:04:38 2022
Response via : Initial Calibration

