

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072222\
 Data File : VU049851.D
 Acq On : 22 Jul 2022 15:45
 Operator : SY/MD
 Sample : N3825-01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GBHY3

Quant Time: Jul 23 02:35:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR072222WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Jul 23 01:04:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	101037	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	101079	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.813	152	45410	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.594	65	35260	4.311	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	86.200%
7) Chloroethane-d5	1.910	69	31607	4.617	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	92.400%
11) 1,1-Dichloroethene-d2	2.562	65	15206	4.165	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	83.200%
20) 2-Butanone-d5	4.633	46	138829	47.170	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	94.340%
24) Chloroform-d	5.064	84	84621	4.645	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.800%
26) 1,2-Dichloroethane-d4	5.704	65	49734	4.656	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.200%
32) Benzene-d6	5.726	84	154764	4.493	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.800%
36) 1,2-Dichloropropane-d6	6.691	67	54555	4.585	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	91.600%
41) Toluene-d8	7.896	98	124759	4.144	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.800%
43) trans-1,3-Dichloroprop...	8.183	79	18550	4.294	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	85.800%
46) 2-Hexanone-d5	8.636	63	77956	45.537	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	91.080%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	45148	4.518	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	90.400%
66) 1,2-Dichlorobenzene-d4	12.195	152	44415	4.739	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	94.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072222\
Data File : VU049851.D
Acq On : 22 Jul 2022 15:45
Operator : SY/MD
Sample : N3825-01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GBHY3

Quant Time: Jul 23 02:35:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR072222WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jul 23 01:04:38 2022
Response via : Initial Calibration

