

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU072319\
 Data File : VU033369.D
 Acq On : 23 Jul 2019 09:34
 Operator : JC/SP
 Sample : VU0723WBL01
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU0723WBL01

Quant Time: Jul 24 02:47:35 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA U\METHOD\524U072219DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Tue Jul 23 03:43:51 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Fluorobenzene	5.72	96	29866	1.00	ug/l	0.00

System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.30	95	9435	0.80	ug/l	0.00
Spiked Amount 1.000			Recovery	=	80.00%	
68) 1,2-Dichlorobenzene-d4	11.85	152	10020	0.83	ug/l	0.00
Spiked Amount 1.000			Recovery	=	83.00%	

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU072319\
Data File : VU033369.D
Acq On : 23 Jul 2019 09:34
Operator : JC/SP
Sample : VU0723WBL01
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0723WBL01

Quant Time: Jul 24 02:47:35 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA U\METHOD\524U072219DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Jul 23 03:43:51 2019
Response via : Initial Calibration

