

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU072319\
 Data File : VU033375.D
 Acq On : 23 Jul 2019 12:22
 Operator : JC/SP
 Sample : K3860-10
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PT-ADD-WS

Quant Time: Jul 25 11:37:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA U\METHOD\524U072219DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Tue Jul 23 03:43:51 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
-----	-----	-----	-----	-----	-----	-----
1) Fluorobenzene	5.72	96	32177	1.00	ug/l	0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.29	95	8925	0.71	ug/l	0.00
Spiked Amount 1.000			Recovery	=	71.00%	
68) 1,2-Dichlorobenzene-d4	11.85	152	11084	0.85	ug/l	0.00
Spiked Amount 1.000			Recovery	=	85.00%	
Target Compounds						
					Ovalue	
8) 1,1,2-Trichloro-1,2,2-trif	2.27	101	448066	38.596	ug/l	99
24) tert-Butyl Alcohol	2.82	59	27757	18.561	ug/l #	73
25) Methyl tert-Butyl Ether	2.98	73	744540	22.616	ug/l	98
31) Isopropyl Ether	3.56	45	631147	27.403	ug/l	98
32) Ethyl-t-butyl ether	4.05	59	1003005	45.299	ug/l	100
33) Tert-Amyl methyl ether	5.56	73	1051438	52.804	ug/l	99
73) n-propylbenzene	10.57	120	461996	40.714	ug/l	92
-----	-----	-----	-----	-----	-----	-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU072319\
Data File : VU033375.D
Acq On : 23 Jul 2019 12:22
Operator : JC/SP
Sample : K3860-10
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PT-ADD-WS

Quant Time: Jul 25 11:37:23 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA U\METHOD\524U072219DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Jul 23 03:43:51 2019
Response via : Initial Calibration

