

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU072319\
 Data File : VU033378.D
 Acq On : 23 Jul 2019 13:43
 Operator : JC/SP
 Sample : K3860-09
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PT-THM-WS

Quant Time: Jul 25 11:41:04 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA U\METHOD\524U072219DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Tue Jul 23 03:43:51 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
-----	-----	-----	-----	-----	-----	-----
1) Fluorobenzene	5.72	96	25068	1.00	ug/l	0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.29	95	8173	0.83	ug/l	0.00
Spiked Amount 1.000			Recovery	=	83.00%	
68) 1,2-Dichlorobenzene-d4	11.85	152	9038	0.89	ug/l	0.00
Spiked Amount 1.000			Recovery	=	89.00%	
Target Compounds						Ovalue
27) Chloroform	4.65	83	789516	48.661	ug/l	100
44) Bromodichloromethane	6.74	83	537666	48.769	ug/l	99
55) Dibromochloromethane	8.46	129	243929	34.172	ug/l	99
67) Bromoform	9.94	173	169104	43.772	ug/l	100
-----	-----	-----	-----	-----	-----	-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU072319\
Data File : VU033378.D
Acq On : 23 Jul 2019 13:43
Operator : JC/SP
Sample : K3860-09
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PT-THM-WS

Quant Time: Jul 25 11:41:04 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA U\METHOD\524U072219DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Jul 23 03:43:51 2019
Response via : Initial Calibration

