

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU072319\
 Data File : VU033382.D
 Acq On : 23 Jul 2019 15:31
 Operator : JC/SP
 Sample : K3860-09DL 4X
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PT-THM-WSDL

Quant Time: Jul 24 04:43:19 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA U\METHOD\524U072219DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Tue Jul 23 03:43:51 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
-----	-----	-----	-----	-----	-----	-----
1) Fluorobenzene	5.72	96	25292	1.00	ug/l	0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.30	95	8209	0.83	ug/l	0.00
Spiked Amount 1.000			Recovery	=	83.00%	
68) 1,2-Dichlorobenzene-d4	11.85	152	8478	0.83	ug/l	0.00
Spiked Amount 1.000			Recovery	=	83.00%	
Target Compounds						Ovalue
27) Chloroform	4.65	83	197468	12.063	ug/l	99
44) Bromodichloromethane	6.74	83	130402	11.723	ug/l	99
55) Dibromochloromethane	8.46	129	56948	7.907	ug/l	98
67) Bromoform	9.94	173	38694	9.927	ug/l	97
-----	-----	-----	-----	-----	-----	-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU072319\
Data File : VU033382.D
Acq On : 23 Jul 2019 15:31
Operator : JC/SP
Sample : K3860-09DL 4X
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PT-THM-WSDL

Quant Time: Jul 24 04:43:19 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA U\METHOD\524U072219DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Jul 23 03:43:51 2019
Response via : Initial Calibration

