

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072324\
 Data File : VU060106.D
 Acq On : 23 Jul 2024 11:47
 Operator : MD/SY
 Sample : VU0723WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK024

Quant Time: Jul 24 04:20:00 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR071824WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Jul 24 04:18:44 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	151199	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	160480	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	66250	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	51366	4.883	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	97.600%
7) Chloroethane-d5	1.908	69	45586	4.664	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	93.200%
11) 1,1-Dichloroethene-d2	2.560	65	19872	4.838	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	96.800%
20) 2-Butanone-d5	4.628	46	97270	39.549	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	79.100%
24) Chloroform-d	5.055	84	92932	4.203	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.000%
26) 1,2-Dichloroethane-d4	5.695	65	45273	4.398	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.000%
32) Benzene-d6	5.721	84	186665	4.341	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.800%
36) 1,2-Dichloropropane-d6	6.685	67	59263	4.325	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	86.600%
41) Toluene-d8	7.894	98	153877	3.961	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	79.200%
43) trans-1,3-Dichloroprop...	8.177	79	18568	3.981	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	79.600%
46) 2-Hexanone-d5	8.631	63	74824	36.154	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	72.300%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	50616	4.151	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	83.000%
66) 1,2-Dichlorobenzene-d4	12.190	152	56485	4.723	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	94.400%
Target Compounds					Qvalue	
16) Methylene chloride	3.039	84	4012	0.284	ug/L	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072324\
Data File : VU060106.D
Acq On : 23 Jul 2024 11:47
Operator : MD/SY
Sample : VU0723WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK024

Quant Time: Jul 24 04:20:00 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR071824WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Jul 24 04:18:44 2024
Response via : Initial Calibration

