

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072324\
 Data File : VU060134.D
 Acq On : 24 Jul 2024 00:27
 Operator : MD/SY
 Sample : P3300-24
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 E1GD4

Quant Time: Jul 24 05:58:06 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR071824WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Jul 24 05:34:19 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	163340	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	167175	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	76060	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	54153	4.765	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	95.400%
7) Chloroethane-d5	1.911	69	46570	4.411	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	88.200%
11) 1,1-Dichloroethene-d2	2.560	65	21056	4.745	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	95.000%
20) 2-Butanone-d5	4.627	46	116691	43.919	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	87.840%
24) Chloroform-d	5.055	84	95527	4.000	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	80.000%
26) 1,2-Dichloroethane-d4	5.695	65	47203	4.244	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.800%
32) Benzene-d6	5.721	84	199986	4.465	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.200%
36) 1,2-Dichloropropane-d6	6.682	67	62399	4.372	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	87.400%
41) Toluene-d8	7.891	98	170359	4.210	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.200%
43) trans-1,3-Dichloroprop...	8.177	79	20313	4.181	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	83.600%
46) 2-Hexanone-d5	8.627	63	89161	41.356	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	82.720%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	54287	4.274	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	85.400%
66) 1,2-Dichlorobenzene-d4	12.187	152	66937	4.876	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	97.600%
Target Compounds						
					Qvalue	
13) Acetone	2.644	43	91415	52.012	ug/L	98
16) Methylene chloride	3.042	84	2996	0.196	ug/L	94
42) Toluene	7.965	91	48722	0.829	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072324\
Data File : VU060134.D
Acq On : 24 Jul 2024 00:27
Operator : MD/SY
Sample : P3300-24
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
E1GD4

Quant Time: Jul 24 05:58:06 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR071824WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Jul 24 05:34:19 2024
Response via : Initial Calibration

