

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072324\
 Data File : VU060143.D
 Acq On : 24 Jul 2024 04:01
 Operator : MD/SY
 Sample : P3300-26
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 E1GE9

Quant Time: Jul 24 06:00:25 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR071824WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Jul 24 05:34:19 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	153637	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	162964	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	70860	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	48667	4.553	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	91.000%	
7) Chloroethane-d5	1.911	69	43718	4.402	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	88.000%	
11) 1,1-Dichloroethene-d2	2.560	65	17969	4.305	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	86.200%	
20) 2-Butanone-d5	4.627	46	99826	39.944	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	79.880%	
24) Chloroform-d	5.055	84	89673	3.992	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	79.800%	
26) 1,2-Dichloroethane-d4	5.695	65	42397	4.053	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	81.000%	
32) Benzene-d6	5.721	84	184327	4.222	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	84.400%	
36) 1,2-Dichloropropane-d6	6.682	67	56158	4.036	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	80.800%	
41) Toluene-d8	7.894	98	155862	3.951	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	79.000%	
43) trans-1,3-Dichloroprop...	8.177	79	16686	3.523	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	70.400%	
46) 2-Hexanone-d5	8.630	63	77638	36.942	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	73.880%	
56) 1,1,2,2-Tetrachloroeth...	10.749	84	49193	3.973	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	79.400%	
66) 1,2-Dichlorobenzene-d4	12.187	152	60595	4.738	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	94.800%	
Target Compounds						
					Qvalue	
13) Acetone	2.644	43	112157	67.843	ug/L	100
16) Methylene chloride	3.039	84	2509	0.175	ug/L	84
42) Toluene	7.965	91	40847	0.713	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072324\
Data File : VU060143.D
Acq On : 24 Jul 2024 04:01
Operator : MD/SY
Sample : P3300-26
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
E1GE9

Quant Time: Jul 24 06:00:25 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR071824WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Jul 24 05:34:19 2024
Response via : Initial Calibration

