




Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072324\
 Data File : VU060117.D
 Acq On : 23 Jul 2024 16:10
 Operator : MD/SY
 Sample : P3300-01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 E1G91

Manual IntegrationsAPPROVED

Quant Time: Jul 24 04:22:43 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR071824WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Jul 24 04:18:44 2024
 Response via : Initial Calibration

Reviewed By :Mahesh Dadoda 07/26/2024
 Supervised By :Semsettin Yesilyurt 07/26/2024



TIC: VU060117.D\data.ms

(13) Acetone (T)

2.447min (-0.196) 595.96 ug/L

response 697495

Ion	Exp%	Act%
43.05	100.00	100.00
58.05	41.10	1.34
0.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072324\
 Data File : VU060117.D
 Acq On : 23 Jul 2024 16:10
 Operator : MD/SY
 Sample : P3300-01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 E1G91

Manual IntegrationsAPPROVED

Quant Time: Jul 24 04:22:43 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR071824WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Jul 24 04:18:44 2024
 Response via : Initial Calibration

Reviewed By :Mahesh Dadoda 07/26/2024
 Supervised By :Semsettin Yesilyurt 07/26/2024



TIC: VU060117.D\data.ms

(13) Acetone (T)

2.637min (-0.006) 19.63 ug/L m

response 22980

Ion	Exp%	Act%
43.05	100.00	100.00
58.05	41.10	40.53
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072324\
 Data File : VU060117.D
 Acq On : 23 Jul 2024 16:10
 Operator : MD/SY
 Sample : P3300-01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 E1G91

Manual IntegrationsAPPROVED

Quant Time: Jul 24 04:22:43 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR071824WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Jul 24 04:18:44 2024
 Response via : Initial Calibration

Reviewed By :Mahesh Dadoda 07/26/2024
 Supervised By :Semsettin Yesilyurt 07/26/2024

Compound	R.T.	Q	Ion	Response	Conc	Units	Dev(Min)

Internal Standards							
1) 1,4-Difluorobenzene	6.245	114		108768	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117		118847	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152		49115	5.000	ug/L	0.00
System Monitoring Compounds							
4) Vinyl Chloride-d3	1.595	65		33467	4.422	ug/L	0.00
Spiked Amount 5.000	Range 40	- 130		Recovery	=	88.400%	
7) Chloroethane-d5	1.907	69		34798	4.950	ug/L	0.00
Spiked Amount 5.000	Range 65	- 130		Recovery	=	99.000%	
11) 1,1-Dichloroethene-d2	2.560	65		13202	4.468	ug/L	0.00
Spiked Amount 5.000	Range 60	- 125		Recovery	=	89.400%	
20) 2-Butanone-d5	4.624	46		99732	56.369	ug/L	0.00
Spiked Amount 50.000	Range 40	- 130		Recovery	=	112.740%	
24) Chloroform-d	5.055	84		79206	4.980	ug/L	0.00
Spiked Amount 5.000	Range 70	- 125		Recovery	=	99.600%	
26) 1,2-Dichloroethane-d4	5.695	65		40384	5.453	ug/L	0.00
Spiked Amount 5.000	Range 70	- 130		Recovery	=	109.000%	
32) Benzene-d6	5.721	84		152616	4.793	ug/L	0.00
Spiked Amount 5.000	Range 70	- 125		Recovery	=	95.800%	
36) 1,2-Dichloropropane-d6	6.685	67		51641	5.090	ug/L	0.00
Spiked Amount 5.000	Range 60	- 140		Recovery	=	101.800%	
41) Toluene-d8	7.894	98		119018	4.137	ug/L	0.00
Spiked Amount 5.000	Range 70	- 130		Recovery	=	82.800%	
43) trans-1,3-Dichloroprop...	8.177	79		9832	2.847	ug/L	0.00
Spiked Amount 5.000	Range 55	- 130		Recovery	=	57.000%	
46) 2-Hexanone-d5	8.627	63		66934	43.671	ug/L	0.00
Spiked Amount 50.000	Range 45	- 130		Recovery	=	87.340%	
56) 1,1,2,2-Tetrachloroeth...	10.749	84		47140	5.220	ug/L	0.00
Spiked Amount 5.000	Range 65	- 120		Recovery	=	104.400%	
66) 1,2-Dichlorobenzene-d4	12.190	152		50529	5.700	ug/L	0.00
Spiked Amount 5.000	Range 80	- 120		Recovery	=	114.000%	
Target Compounds							
13) Acetone	2.637	43		22980m	19.635	ug/L	
42) Toluene	7.965	91		468283	11.208	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072324\
Data File : VU060117.D
Acq On : 23 Jul 2024 16:10
Operator : MD/SY
Sample : P3300-01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
E1G91

Manual IntegrationsAPPROVED

Reviewed By :Mahesh Dadoda 07/26/2024
Supervised By :Semsettin Yesilyurt 07/26/2024

Quant Time: Jul 24 04:22:43 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR071824WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Jul 24 04:18:44 2024
Response via : Initial Calibration

