




Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072324\
 Data File : VU060119.D
 Acq On : 23 Jul 2024 16:58
 Operator : MD/SY
 Sample : P3300-03
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 E1GD3

Manual IntegrationsAPPROVED

Quant Time: Jul 24 04:23:15 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR071824WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Jul 24 04:18:44 2024
 Response via : Initial Calibration

Reviewed By :Mahesh Dadoda 07/26/2024
 Supervised By :Semsettin Yesilyurt 07/26/2024



TIC: VU060119.D\data.ms

(13) Acetone (T)

2.448min (-0.196) 312.26 ug/L

response 375312

Ion	Exp%	Act%
43.05	100.00	100.00
58.05	41.10	3.03
0.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072324\
Data File : VU060119.D
Acq On : 23 Jul 2024 16:58
Operator : MD/SY
Sample : P3300-03
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
E1GD3

Manual IntegrationsAPPROVED

Quant Time: Jul 24 04:23:15 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR071824WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Jul 24 04:18:44 2024
Response via : Initial Calibration

Reviewed By :Mahesh Dadoda 07/26/2024
Supervised By :Semsettin Yesilyurt 07/26/2024



TIC: VU060119.D\data.ms

(13) Acetone (T)

2.641min (-0.003) 24.86 ug/L m

response 29877

Ion	Exp%	Act%
43.05	100.00	100.00
58.05	41.10	38.12
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072324\
 Data File : VU060119.D
 Acq On : 23 Jul 2024 16:58
 Operator : MD/SY
 Sample : P3300-03
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 E1GD3

Manual IntegrationsAPPROVED

Quant Time: Jul 24 04:23:15 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR071824WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Jul 24 04:18:44 2024
 Response via : Initial Calibration

Reviewed By :Mahesh Dadoda 07/26/2024
 Supervised By :Semsettin Yesilyurt 07/26/2024

Compound	R.T.	Q	Ion	Response	Conc	Units	Dev(Min)

Internal Standards							
1) 1,4-Difluorobenzene	6.245	114		111698	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117		119711	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152		50351	5.000	ug/L	0.00
System Monitoring Compounds							
4) Vinyl Chloride-d3	1.596	65		33934	4.366	ug/L	0.00
Spiked Amount 5.000	Range 40	- 130		Recovery =	87.400%		
7) Chloroethane-d5	1.907	69		35263	4.884	ug/L	0.00
Spiked Amount 5.000	Range 65	- 130		Recovery =	97.600%		
11) 1,1-Dichloroethene-d2	2.560	65		13340	4.396	ug/L	0.00
Spiked Amount 5.000	Range 60	- 125		Recovery =	88.000%		
20) 2-Butanone-d5	4.628	46		96699	53.221	ug/L	0.00
Spiked Amount 50.000	Range 40	- 130		Recovery =	106.440%		
24) Chloroform-d	5.055	84		81553	4.993	ug/L	0.00
Spiked Amount 5.000	Range 70	- 125		Recovery =	99.800%		
26) 1,2-Dichloroethane-d4	5.695	65		40852	5.372	ug/L	0.00
Spiked Amount 5.000	Range 70	- 130		Recovery =	107.400%		
32) Benzene-d6	5.721	84		156119	4.868	ug/L	0.00
Spiked Amount 5.000	Range 70	- 125		Recovery =	97.400%		
36) 1,2-Dichloropropane-d6	6.685	67		52002	5.088	ug/L	0.00
Spiked Amount 5.000	Range 60	- 140		Recovery =	101.800%		
41) Toluene-d8	7.894	98		122137	4.215	ug/L	0.00
Spiked Amount 5.000	Range 70	- 130		Recovery =	84.200%		
43) trans-1,3-Dichloroprop...	8.177	79		8833	2.539	ug/L	0.00
Spiked Amount 5.000	Range 55	- 130		Recovery =	50.800%#		
46) 2-Hexanone-d5	8.627	63		58878	38.138	ug/L	0.00
Spiked Amount 50.000	Range 45	- 130		Recovery =	76.280%		
56) 1,1,2,2-Tetrachloroeth...	10.750	84		48506	5.333	ug/L	0.00
Spiked Amount 5.000	Range 65	- 120		Recovery =	106.600%		
66) 1,2-Dichlorobenzene-d4	12.190	152		52861	5.816	ug/L	0.00
Spiked Amount 5.000	Range 80	- 120		Recovery =	116.400%		
Target Compounds							
13) Acetone	2.641	43		29877m	24.858	ug/L	
16) Methylene chloride	3.033	84		2777	0.266	ug/L	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072324\
Data File : VU060119.D
Acq On : 23 Jul 2024 16:58
Operator : MD/SY
Sample : P3300-03
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
E1GD3

Manual IntegrationsAPPROVED

Quant Time: Jul 24 04:23:15 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR071824WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Jul 24 04:18:44 2024
Response via : Initial Calibration

Reviewed By :Mahesh Dadoda 07/26/2024
Supervised By :Semsettin Yesilyurt 07/26/2024

