

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072423\
 Data File : VU054850.D
 Acq On : 24 Jul 2023 10:58
 Operator : MD/SY
 Sample : VU0724WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU0724WBL01

Quant Time: Jul 25 02:12:37 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U072123DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Mon Jul 24 09:59:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Fluorobenzene	6.113	96	26685	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.631	95	7421	0.921	ug/l	0.00
Spiked Amount 1.000			Recovery	=	92.000%	
68) 1,2-Dichlorobenzene-d4	12.196	152	9071	0.972	ug/l	0.00
Spiked Amount 1.000			Recovery	=	97.000%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072423\
Data File : VU054850.D
Acq On : 24 Jul 2023 10:58
Operator : MD/SY
Sample : VU0724WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0724WBL01

Quant Time: Jul 25 02:12:37 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U072123DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Mon Jul 24 09:59:15 2023
Response via : Initial Calibration

