

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072423\
 Data File : VU054858.D
 Acq On : 24 Jul 2023 14:39
 Operator : MD/SY
 Sample : 03587-11
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PT-EDBCP-WS

Quant Time: Jul 25 02:14:41 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U072123DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Mon Jul 24 09:59:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Fluorobenzene	6.113	96	25232	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.631	95	7752	1.017	ug/l	0.00
Spiked Amount 1.000			Recovery	=	102.000%	
68) 1,2-Dichlorobenzene-d4	12.193	152	8649	0.980	ug/l	0.00
Spiked Amount 1.000			Recovery	=	98.000%	
Target Compounds						
15) Methylene Chloride	3.039	84	5272	0.246	ug/l	97
71) 1,2,3-Trichloropropane	10.823	75	2427	0.626	ug/l #	91
86) 1,2-Dibromo-3-Chloropr...	12.997	75	1221	1.815	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072423\
Data File : VU054858.D
Acq On : 24 Jul 2023 14:39
Operator : MD/SY
Sample : 03587-11
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PT-EDBCP-WS

Quant Time: Jul 25 02:14:41 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U072123DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Mon Jul 24 09:59:15 2023
Response via : Initial Calibration

