

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072423\
 Data File : VU054862.D
 Acq On : 24 Jul 2023 16:17
 Operator : MD/SY
 Sample : 03587-09
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PT-THM-WS

Quant Time: Jul 25 05:15:50 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U072123DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Tue Jul 25 05:02:28 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Fluorobenzene	6.110	96	27488	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.631	95	7681	0.925	ug/l	0.00
Spiked Amount 1.000			Recovery	=	93.000%	
68) 1,2-Dichlorobenzene-d4	12.193	152	8167	0.850	ug/l	0.00
Spiked Amount 1.000			Recovery	=	85.000%	
Target Compounds						
						Qvalue
15) Methylene Chloride	3.033	84	2415	0.274	ug/l	# 77
27) Chloroform	5.081	83	292616	23.948	ug/l	95
44) Bromodichloromethane	7.100	83	304732	34.897	ug/l	# 98
55) Dibromochloromethane	8.807	129	68075	11.848	ug/l	98
67) Bromoform	10.287	173	122027	39.414	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072423\
Data File : VU054862.D
Acq On : 24 Jul 2023 16:17
Operator : MD/SY
Sample : 03587-09
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PT-THM-WS

Quant Time: Jul 25 05:15:50 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U072123DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Jul 25 05:02:28 2023
Response via : Initial Calibration

