

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU072519\
 Data File : VU033426.D
 Acq On : 25 Jul 2019 21:51
 Operator : JC/SP
 Sample : VIBLK40
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK40

Quant Time: Jul 26 07:14:16 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM071219WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jul 26 05:48:12 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.86	114	266460	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.07	117	276553	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.47	152	117472	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	82962	41.13	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	82.26%
7) Chloroethane-d5	1.67	69	82339	47.78	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	95.56%
11) 1,1-Dichloroethene-d2	2.25	63	120866	33.03	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	66.06%
21) 2-Butanone-d5	4.15	46	133467	104.43	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	104.43%
24) Chloroform-d	4.62	84	160440	47.46	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.92%
26) 1,2-Dichloroethane-d4	5.28	65	110707	51.06	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.12%
32) Benzene-d6	5.32	84	318011	46.87	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.74%
36) 1,2-Dichloropropane-d6	6.31	67	107703	49.56	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	99.12%
41) Toluene-d8	7.55	98	281075	43.93	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	87.86%
43) trans-1,3-Dichloropropene-	7.83	79	46220	43.37	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	86.74%
47) 2-Hexanone-d5	8.29	63	95011	96.98	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	96.98%
57) 1,1,2,2-Tetrachloroethane-	10.41	84	169090	51.41	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	102.82%
64) 1,2-Dichlorobenzene-d4	11.85	152	113684	54.27	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	108.54%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU072519\
Data File : VU033426.D
Acq On : 25 Jul 2019 21:51
Operator : JC/SP
Sample : VIBLK40
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK40

Quant Time: Jul 26 07:14:16 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM071219WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jul 26 05:48:12 2019
Response via : Initial Calibration

