

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072522\
 Data File : VU049872.D
 Acq On : 25 Jul 2022 17:16
 Operator : SY/MD
 Sample : N3802-16
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GBHX8

Quant Time: Jul 26 05:35:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR072222WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Jul 26 05:33:39 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	87485	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	81895	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.816	152	34733	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	36188	5.109	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 102.200%		
7) Chloroethane-d5	1.909	69	31522	5.318	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	= 106.400%		
11) 1,1-Dichloroethene-d2	2.565	65	15151	4.793	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 95.800%		
20) 2-Butanone-d5	4.629	46	134619	52.825	ug/L	-0.02
Spiked Amount	50.000	Range 40 - 130	Recovery	= 105.660%		
24) Chloroform-d	5.067	84	81402	5.160	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 103.200%		
26) 1,2-Dichloroethane-d4	5.703	65	46070	4.981	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 99.600%		
32) Benzene-d6	5.729	84	147977	5.303	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 106.000%		
36) 1,2-Dichloropropane-d6	6.690	67	52836	5.480	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 109.600%		
41) Toluene-d8	7.899	98	118973	4.877	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 97.600%		
43) trans-1,3-Dichloroprop...	8.182	79	17086	4.881	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 97.600%		
46) 2-Hexanone-d5	8.636	63	68833	49.627	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 99.260%		
56) 1,1,2,2-Tetrachloroeth...	10.758	84	42213	5.214	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 104.200%		
66) 1,2-Dichlorobenzene-d4	12.195	152	40654	5.671	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 113.400%		
Target Compounds						
34) Trichloroethene	6.542	95	26368	3.660	ug/L	98
47) Tetrachloroethene	8.552	164	2456	0.495	ug/L	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072522\
Data File : VU049872.D
Acq On : 25 Jul 2022 17:16
Operator : SY/MD
Sample : N3802-16
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GBHX8

Quant Time: Jul 26 05:35:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR072222WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Jul 26 05:33:39 2022
Response via : Initial Calibration

