

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072522\
 Data File : VU049882.D
 Acq On : 25 Jul 2022 22:09
 Operator : SY/MD
 Sample : N3879-01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DBUF2

Quant Time: Jul 26 06:47:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR072222WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Jul 26 06:46:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.263	114	79559	5.000	ug/L	0.01
28) Chlorobenzene-d5	9.420	117	77068	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.815	152	31655	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	32745	5.084	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	101.600%
7) Chloroethane-d5	1.912	69	29377	5.450	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	109.000%
11) 1,1-Dichloroethene-d2	2.571	65	14685	5.108	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	102.200%
20) 2-Butanone-d5	4.671	46	111675	48.187	ug/L	0.03
Spiked Amount	50.000	Range	40 - 130	Recovery	=	96.380%
24) Chloroform-d	5.070	84	74258	5.176	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.600%
26) 1,2-Dichloroethane-d4	5.726	65	42877	5.097	ug/L	0.02
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.000%
32) Benzene-d6	5.742	84	133490	5.083	ug/L	0.02
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.600%
36) 1,2-Dichloropropane-d6	6.703	67	48007	5.291	ug/L	0.01
Spiked Amount	5.000	Range	60 - 140	Recovery	=	105.800%
41) Toluene-d8	7.906	98	106598	4.644	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.800%
43) trans-1,3-Dichloroprop...	8.189	79	15649	4.751	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	95.000%
46) 2-Hexanone-d5	8.648	63	67778	51.927	ug/L	0.01
Spiked Amount	50.000	Range	45 - 130	Recovery	=	103.860%
56) 1,1,2,2-Tetrachloroeth...	10.761	84	41483	5.445	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	108.800%
66) 1,2-Dichlorobenzene-d4	12.195	152	36736	5.622	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	112.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072522\
Data File : VU049882.D
Acq On : 25 Jul 2022 22:09
Operator : SY/MD
Sample : N3879-01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DBUF2

Quant Time: Jul 26 06:47:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR072222WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Jul 26 06:46:04 2022
Response via : Initial Calibration

