


Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072523\
Data File : VU054865.D
Acq On : 25 Jul 2023 08:52
Operator : MD/SY
Sample : BFB
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFB

Integration File: rteint.p

Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U072523DW.M
Title : METHOD 524.2 VOLATILES DRINKING WATER
Last Update : Wed Jul 26 09:45:35 2023



AutoFind: Scans 2904, 2905, 2906; Background Corrected with Scan 2892

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.3	14800	PASS
75	95	30	60	58.1	38541	PASS
95	95	100	100	100.0	66347	PASS
96	95	5	9	7.1	4690	PASS
173	174	0.00	2	1.1	643	PASS
174	95	50	100	90.3	59885	PASS
175	174	5	9	7.5	4463	PASS
176	174	95	101	96.8	57955	PASS
177	176	5	9	6.4	3705	PASS