

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU072618\
 Data File : VU025649.D
 Acq On : 26 Jul 2018 14:37
 Operator : MD/SY
 Sample : J4148-01
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0SW6

Quant Time: Jul 27 02:35:29 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM072018WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jul 27 02:03:29 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	308717	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	290500	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.49	152	152893	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	76955	43.66	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	87.32%
7) Chloroethane-d5	1.68	69	66355	45.64	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	91.28%
11) 1,1-Dichloroethene-d2	2.27	63	118596	32.32	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	64.64%
21) 2-Butanone-d5	4.18	46	137469	101.03	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	101.03%
24) Chloroform-d	4.65	84	174329	44.41	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.82%
26) 1,2-Dichloroethane-d4	5.31	65	123989	47.76	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.52%
32) Benzene-d6	5.35	84	341874	45.89	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.78%
36) 1,2-Dichloropropane-d6	6.33	67	110892	46.19	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	92.38%
41) Toluene-d8	7.57	98	317750	44.47	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	88.94%
43) trans-1,3-Dichloropropene-	7.85	79	53414	43.56	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	87.12%
47) 2-Hexanone-d5	8.31	63	96089	95.59	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	95.59%
57) 1,1,2,2-Tetrachloroethane-	10.44	84	150930	44.95	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	89.90%
64) 1,2-Dichlorobenzene-d4	11.86	152	138752	44.78	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	89.56%

Target Compounds

					Ovalue
9) Trichlorofluoromethane	1.89	101	32826	9.567 ug/L	98
20) cis-1,2-Dichloroethene	4.23	96	5368	2.371 ug/L	98
34) Trichloroethene	6.19	95	248856	112.365 ug/L	97
46) Tetrachloroethene	8.23	164	85936	44.758 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU072618\
Data File : VU025649.D
Acq On : 26 Jul 2018 14:37
Operator : MD/SY
Sample : J4148-01
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0SW6

Quant Time: Jul 27 02:35:29 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM072018WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jul 27 02:03:29 2018
Response via : Initial Calibration

