

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU072618\
 Data File : VU025650.D
 Acq On : 26 Jul 2018 15:00
 Operator : MD/SY
 Sample : J4148-05
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0SX0

Quant Time: Jul 27 02:36:41 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM072018WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jul 27 02:03:29 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	300430	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	286528	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.49	152	148445	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	73956	43.12	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	86.24%
7) Chloroethane-d5	1.68	69	63257	44.71	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	89.42%
11) 1,1-Dichloroethene-d2	2.27	63	111787	31.31	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	62.62%
21) 2-Butanone-d5	4.18	46	136057	102.75	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	102.75%
24) Chloroform-d	4.65	84	167789	43.92	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.84%
26) 1,2-Dichloroethane-d4	5.31	65	117857	46.65	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.30%
32) Benzene-d6	5.35	84	323167	43.98	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.96%
36) 1,2-Dichloropropane-d6	6.34	67	107718	45.49	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	90.98%
41) Toluene-d8	7.57	98	301294	42.75	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	85.50%
43) trans-1,3-Dichloropropene-	7.85	79	51130	42.27	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	84.54%
47) 2-Hexanone-d5	8.31	63	95236	96.06	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	96.06%
57) 1,1,2,2-Tetrachloroethane-	10.44	84	148316	44.79	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	89.58%
64) 1,2-Dichlorobenzene-d4	11.86	152	131912	43.85	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	87.70%

Target Compounds

					Ovalue
9) Trichlorofluoromethane	1.89	101	26477	7.930 ug/L	99
20) cis-1,2-Dichloroethene	4.23	96	5191	2.356 ug/L	100
34) Trichloroethene	6.19	95	305757	139.971 ug/L	95
46) Tetrachloroethene	8.23	164	70502	37.229 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU072618\
Data File : VU025650.D
Acq On : 26 Jul 2018 15:00
Operator : MD/SY
Sample : J4148-05
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0SX0

Quant Time: Jul 27 02:36:41 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM072018WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jul 27 02:03:29 2018
Response via : Initial Calibration

