

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072622\
 Data File : VU049907.D
 Acq On : 26 Jul 2022 11:53
 Operator : SY/MD
 Sample : VU0726WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK098

Quant Time: Jul 27 03:56:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR072222WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Jul 27 03:55:15 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	91609	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	90927	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	38635	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	31683	4.272	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	85.400%
7) Chloroethane-d5	1.912	69	29001	4.672	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	93.400%
11) 1,1-Dichloroethene-d2	2.565	65	13785	4.164	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	83.200%
20) 2-Butanone-d5	4.642	46	133635	50.078	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	100.160%
24) Chloroform-d	5.063	84	69319	4.197	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.000%
26) 1,2-Dichloroethane-d4	5.703	65	41676	4.303	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.000%
32) Benzene-d6	5.729	84	121561	3.923	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	78.400%
36) 1,2-Dichloropropane-d6	6.690	67	45454	4.246	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	85.000%
41) Toluene-d8	7.899	98	95159	3.514	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	70.200%
43) trans-1,3-Dichloroprop...	8.182	79	14986	3.856	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	77.200%
46) 2-Hexanone-d5	8.635	63	62713	40.723	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	81.440%
56) 1,1,2,2-Tetrachloroeth...	10.754	84	38953	4.333	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	86.600%
66) 1,2-Dichlorobenzene-d4	12.195	152	33555	4.208	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	84.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072622\
Data File : VU049907.D
Acq On : 26 Jul 2022 11:53
Operator : SY/MD
Sample : VU0726WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK098

Quant Time: Jul 27 03:56:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR072222WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Jul 27 03:55:15 2022
Response via : Initial Calibration

