

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072622\
 Data File : VU049917.D
 Acq On : 26 Jul 2022 17:10
 Operator : SY/MD
 Sample : N3870-08DL 20X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GBJ07DL

Quant Time: Jul 27 03:58:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR072222WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Jul 27 03:55:15 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	85840	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	84841	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	38753	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	27487	3.955	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	79.200%
7) Chloroethane-d5	1.912	69	25385	4.365	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	87.200%
11) 1,1-Dichloroethene-d2	2.565	65	12104	3.902	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	78.000%
20) 2-Butanone-d5	4.633	46	127578	51.021	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	102.040%
24) Chloroform-d	5.063	84	68919	4.453	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.000%
26) 1,2-Dichloroethane-d4	5.703	65	40025	4.410	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.200%
32) Benzene-d6	5.729	84	116711	4.037	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	80.800%
36) 1,2-Dichloropropane-d6	6.690	67	44023	4.408	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	88.200%
41) Toluene-d8	7.896	98	91403	3.617	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	72.400%
43) trans-1,3-Dichloroprop...	8.179	79	13908	3.835	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	76.800%
46) 2-Hexanone-d5	8.636	63	63954	44.508	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	89.020%
56) 1,1,2,2-Tetrachloroeth...	10.754	84	40784	4.863	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.200%
66) 1,2-Dichlorobenzene-d4	12.195	152	34241	4.281	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	85.600%
Target Compounds						
					Qvalue	
10) 1,1,2-Trichloro-1,2,2-...	2.578	101	3978	0.710	ug/L	91
12) 1,1-Dichloroethene	2.578	96	1138	0.201	ug/L	# 1
19) 1,1-Dichloroethane	3.874	63	1646	0.116	ug/L	91
22) cis-1,2-Dichloroethene	4.668	96	6274	0.826	ug/L	96
34) Trichloroethene	6.542	95	58979	7.901	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072622\
Data File : VU049917.D
Acq On : 26 Jul 2022 17:10
Operator : SY/MD
Sample : N3870-08DL 20X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GBJ07DL

Quant Time: Jul 27 03:58:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR072222WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Jul 27 03:55:15 2022
Response via : Initial Calibration

