

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072622\
 Data File : VU049929.D
 Acq On : 26 Jul 2022 22:55
 Operator : SY/MD
 Sample : VU0726WBL02
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK099

Quant Time: Jul 27 05:18:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR072222WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Jul 27 05:17:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	83484	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	83622	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	35601	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.594	65	27314	4.041	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	80.800%
7) Chloroethane-d5	1.912	69	26604	4.703	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	94.000%
11) 1,1-Dichloroethene-d2	2.565	65	12184	4.039	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	80.800%
20) 2-Butanone-d5	4.639	46	138498	56.952	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	113.900%
24) Chloroform-d	5.063	84	70275	4.669	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.400%
26) 1,2-Dichloroethane-d4	5.703	65	42278	4.790	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.800%
32) Benzene-d6	5.729	84	123816	4.345	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.000%
36) 1,2-Dichloropropane-d6	6.690	67	46246	4.698	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.000%
41) Toluene-d8	7.899	98	93907	3.770	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	75.400%
43) trans-1,3-Dichloroprop...	8.182	79	14472	4.049	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	81.000%
46) 2-Hexanone-d5	8.636	63	61439	43.381	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	86.760%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	41487	5.018	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	100.400%
66) 1,2-Dichlorobenzene-d4	12.195	152	35150	4.783	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	95.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072622\
Data File : VU049929.D
Acq On : 26 Jul 2022 22:55
Operator : SY/MD
Sample : VU0726WBL02
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK099

Quant Time: Jul 27 05:18:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR072222WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Jul 27 05:17:43 2022
Response via : Initial Calibration

