

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072622\
 Data File : VU049944.D
 Acq On : 27 Jul 2022 05:01
 Operator : SY/MD
 Sample : N3879-16
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DBUJ0

Quant Time: Jul 27 06:47:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR072222WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Jul 27 05:17:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	75184	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	76436	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	34087	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	20009	3.287	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	65.800%
7) Chloroethane-d5	1.912	69	20410	4.007	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	80.200%
11) 1,1-Dichloroethene-d2	2.565	65	8793	3.237	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	64.800%
20) 2-Butanone-d5	4.636	46	127897	58.399	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	116.800%
24) Chloroform-d	5.063	84	58769	4.335	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.800%
26) 1,2-Dichloroethane-d4	5.703	65	36155	4.548	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.000%
32) Benzene-d6	5.729	84	98674	3.788	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	75.800%
36) 1,2-Dichloropropane-d6	6.690	67	38742	4.305	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	86.200%
41) Toluene-d8	7.899	98	74063	3.253	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	65.000%#
43) trans-1,3-Dichloroprop...	8.182	79	11804	3.613	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	72.200%
46) 2-Hexanone-d5	8.636	63	57384	44.327	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	88.660%
56) 1,1,2,2-Tetrachloroeth...	10.754	84	37741	4.995	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	99.800%
66) 1,2-Dichlorobenzene-d4	12.192	152	31008	4.407	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	88.200%
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.385	85	1665	0.210	ug/L	# 82
22) cis-1,2-Dichloroethene	4.671	96	2518	0.379	ug/L	93
34) Trichloroethene	6.542	95	5900	0.877	ug/L	95
47) Tetrachloroethene	8.552	164	23190	5.009	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072622\
Data File : VU049944.D
Acq On : 27 Jul 2022 05:01
Operator : SY/MD
Sample : N3879-16
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DBUJ0

Quant Time: Jul 27 06:47:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR072222WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Jul 27 05:17:43 2022
Response via : Initial Calibration

