

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072622\
 Data File : VU049922.D
 Acq On : 26 Jul 2022 19:13
 Operator : SY/MD
 Sample : N3870-10DL 25X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GBJ09DL

Quant Time: Jul 27 03:59:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR072222WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Jul 27 03:55:15 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	76361	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	77871	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	35037	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	23800	3.850	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	77.000%	
7) Chloroethane-d5	1.913	69	23213	4.487	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	89.800%	
11) 1,1-Dichloroethene-d2	2.565	65	10386	3.764	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	75.200%	
20) 2-Butanone-d5	4.636	46	128700	57.859	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	115.720%	
24) Chloroform-d	5.067	84	67072	4.871	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	97.400%	
26) 1,2-Dichloroethane-d4	5.703	65	39206	4.856	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	97.200%	
32) Benzene-d6	5.729	84	111098	4.187	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	83.800%	
36) 1,2-Dichloropropane-d6	6.690	67	44090	4.809	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	96.200%	
41) Toluene-d8	7.899	98	87443	3.770	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	75.400%	
43) trans-1,3-Dichloroprop...	8.179	79	13254	3.982	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	79.600%	
46) 2-Hexanone-d5	8.636	63	60956	46.219	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	92.440%	
56) 1,1,2,2-Tetrachloroeth...	10.755	84	39861	5.178	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	103.600%	
66) 1,2-Dichlorobenzene-d4	12.195	152	35271	4.877	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	97.600%	
Target Compounds						Qvalue
22) cis-1,2-Dichloroethene	4.665	96	4026	0.596	ug/L	88
34) Trichloroethene	6.543	95	71798	10.480	ug/L	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072622\
Data File : VU049922.D
Acq On : 26 Jul 2022 19:13
Operator : SY/MD
Sample : N3870-10DL 25X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GBJ09DL

Quant Time: Jul 27 03:59:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR072222WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Jul 27 03:55:15 2022
Response via : Initial Calibration





#22

cis-1,2-Dichloroethene

Concen: 0.596 ug/L

RT: 4.665 min Scan# 1034

Delta R.T. -0.003 min

Lab File: VU049922.D

Acq: 26 Jul 2022 19:13

Instrument :

MSVOA_U

ClientSampleId :

GBJ09DL





Tgt Ion: 96 Resp: 4026

Ion Ratio Lower Upper

96 100

61 149.9 93.9 174.3

98 55.5 43.9 81.5



#34

Trichloroethene

Concen: 10.480 ug/L

RT: 6.543 min Scan# 1618

Delta R.T. 0.000 min

Lab File: VU049922.D

Acq: 26 Jul 2022 19:13

Tgt Ion: 95 Resp: 71798

Ion Ratio Lower Upper

95 100

97 64.5 46.8 86.8

132 86.2 61.8 114.8

130 91.7 66.3 123.1

