

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072623\
 Data File : VU054880.D
 Acq On : 26 Jul 2023 16:54
 Operator : MD/SY
 Sample : 03587-09DL 5X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PT-THM-WSDL

Quant Time: Jul 28 03:22:45 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U072523DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Wed Jul 26 09:45:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Fluorobenzene	6.113	96	28065	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.634	95	8154	0.958	ug/l	0.00
Spiked Amount 1.000			Recovery	=	96.000%	
68) 1,2-Dichlorobenzene-d4	12.193	152	9862	1.026	ug/l	0.00
Spiked Amount 1.000			Recovery	=	103.000%	
Target Compounds						
27) Chloroform	5.084	83	56485	4.458	ug/l	97
44) Bromodichloromethane	7.103	83	55441	5.929	ug/l #	97
55) Dibromochloromethane	8.807	129	12487	2.048	ug/l	97
67) Bromoform	10.290	173	21867	6.917	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072623\
Data File : VU054880.D
Acq On : 26 Jul 2023 16:54
Operator : MD/SY
Sample : 03587-09DL 5X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PT-THM-WSDL

Quant Time: Jul 28 03:22:45 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U072523DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Wed Jul 26 09:45:35 2023
Response via : Initial Calibration

