

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU072718\
 Data File : VU025678.D
 Acq On : 27 Jul 2018 18:49
 Operator : MD/SY
 Sample : J3762-08 LOO 10.0PPB
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 LOQ-WATER-02-QT3-2018

Quant Time: Jul 30 04:24:33 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U072718W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 27 13:11:52 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	162451	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	245525	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	236604	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	144091	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	133119	52.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.18%	
35) Dibromofluoromethane	4.89	113	106229	49.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.50%	
50) Toluene-d8	7.57	98	378146	50.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.20%	
62) 4-Bromofluorobenzene	10.31	95	144218	49.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.44%	
Target Compounds						
10) Methyl Iodide	2.41	142	10064	13.43	ug/l	99
11) Tert butyl alcohol	2.83	59	29182	52.66	ug/l	99
13) Acrolein	2.19	56	14004	68.30	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU072718\
Data File : VU025678.D
Acq On : 27 Jul 2018 18:49
Operator : MD/SY
Sample : J3762-08 L00 10.0PPB
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
LOQ-WATER-02-QT3-2018

Quant Time: Jul 30 04:24:33 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U072718W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 27 13:11:52 2018
Response via : Initial Calibration

