

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU072718\
 Data File : VU025702.D
 Acq On : 28 Jul 2018 04:59
 Operator : MD/SY
 Sample : J4069-04
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 KY030SP005-180718

Quant Time: Jul 30 09:00:53 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U072718W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 27 13:11:52 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	170693	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	266206	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	262445	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	157443	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	5.31	65	131494	49.44	ug/l	0.00
Spiked Amount			Recovery	=	98.88%	
35) Dibromofluoromethane	4.89	113	110508	47.73	ug/l	0.00
Spiked Amount			Recovery	=	95.46%	
50) Toluene-d8	7.57	98	393137	48.52	ug/l	0.00
Spiked Amount			Recovery	=	97.04%	
62) 4-Bromofluorobenzene	10.31	95	152660	48.54	ug/l	0.00
Spiked Amount			Recovery	=	97.08%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
44) Trichloroethene	6.19	130	4973	1.96	ug/l	89
60) Dibromochloromethane	8.48	129	1171	0.42	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU072718\
Data File : VU025702.D
Acq On : 28 Jul 2018 04:59
Operator : MD/SY
Sample : J4069-04
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
KY030SP005-180718

Quant Time: Jul 30 09:00:53 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U072718W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 27 13:11:52 2018
Response via : Initial Calibration

