

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU072718\
 Data File : VU025708.D
 Acq On : 28 Jul 2018 07:18
 Operator : MD/SY
 Sample : J4208-05
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 54 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 FL-0526

Quant Time: Jul 30 09:14:54 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U072718W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 27 13:11:52 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	180840	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	286550	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	279461	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	170500	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	142982	50.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.48%	
35) Dibromofluoromethane	4.89	113	117441	47.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.24%	
50) Toluene-d8	7.57	98	410040	47.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.04%	
62) 4-Bromofluorobenzene	10.31	95	163560	48.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.64%	
Target Compounds						
67) Ethyl Benzene	9.26	91	3995	0.33	ug/l	Qvalue 100
68) m/p-Xylenes	9.38	106	4351	0.95	ug/l	97
84) 1,2,4-Trimethylbenzene	11.15	105	4648	0.46	ug/l	92
95) Naphthalene	13.75	128	8979	1.94	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU072718\
Data File : VU025708.D
Acq On : 28 Jul 2018 07:18
Operator : MD/SY
Sample : J4208-05
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 54 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
FL-0526

Quant Time: Jul 30 09:14:54 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U072718W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 27 13:11:52 2018
Response via : Initial Calibration

