

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072722\
 Data File : VU049967.D
 Acq On : 27 Jul 2022 23:54
 Operator : SY/MD
 Sample : N3870-23
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Jul 28 03:51:13 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR072722WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Jul 28 03:41:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc Units	Dev(Min)

Internal Standards					
1) 1,4-Difluorobenzene	6.250	114	91750	5.000 ug/L	0.00
28) Chlorobenzene-d5	9.417	117	88876	5.000 ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	36283	5.000 ug/L	0.00
System Monitoring Compounds					
4) Vinyl Chloride-d3	1.597	65	40258	5.110 ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 102.200%	
7) Chloroethane-d5	1.912	69	31931	5.643 ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	= 112.800%	
11) 1,1-Dichloroethene-d2	2.565	65	14979	5.517 ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 110.400%	
20) 2-Butanone-d5	4.629	46	97819	55.782 ug/L	-0.03
Spiked Amount	50.000	Range 40 - 130	Recovery	= 111.560%	
24) Chloroform-d	5.063	84	69828	5.357 ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 107.200%	
26) 1,2-Dichloroethane-d4	5.703	65	36664	5.582 ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 111.600%	
32) Benzene-d6	5.729	84	142829	5.439 ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 108.800%	
36) 1,2-Dichloropropane-d6	6.690	67	47754	5.550 ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 111.000%	
41) Toluene-d8	7.899	98	116632	4.996 ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 100.000%	
43) trans-1,3-Dichloroprop...	8.179	79	16175	5.364 ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 107.200%	
46) 2-Hexanone-d5	8.636	63	56643	50.179 ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 100.360%	
56) 1,1,2,2-Tetrachloroeth...	10.754	84	40270	5.571 ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 111.400%	
66) 1,2-Dichlorobenzene-d4	12.192	152	37773	5.613 ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 112.200%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072722\
Data File : VU049967.D
Acq On : 27 Jul 2022 23:54
Operator : SY/MD
Sample : N3870-23
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Jul 28 03:51:13 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR072722WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Jul 28 03:41:29 2022
Response via : Initial Calibration

