

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072722\
 Data File : VU049968.D
 Acq On : 28 Jul 2022 00:18
 Operator : SY/MD
 Sample : N3922-18
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0CB0

Quant Time: Jul 28 03:51:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR072722WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Jul 28 03:41:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	85169	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	82908	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	33046	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.594	65	45880	6.274	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 125.400%		
7) Chloroethane-d5	1.912	69	34203	6.512	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	= 130.200%#		
11) 1,1-Dichloroethene-d2	2.565	65	16408	6.510	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 130.200%#		
20) 2-Butanone-d5	4.636	46	101244	62.196	ug/L	-0.02
Spiked Amount	50.000	Range 40 - 130	Recovery	= 124.400%		
24) Chloroform-d	5.063	84	71436	5.903	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 118.000%		
26) 1,2-Dichloroethane-d4	5.703	65	36638	6.009	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 120.200%		
32) Benzene-d6	5.726	84	149830	6.116	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 122.400%		
36) 1,2-Dichloropropane-d6	6.690	67	48552	6.049	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 121.000%		
41) Toluene-d8	7.899	98	122744	5.637	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 112.800%		
43) trans-1,3-Dichloroprop...	8.179	79	15989	5.684	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 113.600%		
46) 2-Hexanone-d5	8.636	63	56522	53.676	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 107.360%		
56) 1,1,2,2-Tetrachloroeth...	10.758	84	40694	6.035	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 120.600%#		
66) 1,2-Dichlorobenzene-d4	12.195	152	38988	6.361	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 127.200%#		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072722\
Data File : VU049968.D
Acq On : 28 Jul 2022 00:18
Operator : SY/MD
Sample : N3922-18
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0CB0

Quant Time: Jul 28 03:51:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR072722WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Jul 28 03:41:29 2022
Response via : Initial Calibration

