

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072822\
 Data File : VU050002.D
 Acq On : 28 Jul 2022 22:32
 Operator : SY/MD
 Sample : VU0728WBL02
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK102

Quant Time: Jul 29 05:05:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR072722WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Jul 29 05:04:48 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	92012	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	93207	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	36046	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	37546	4.753	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	95.000%	
7) Chloroethane-d5	1.913	69	31164	5.492	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	109.800%	
11) 1,1-Dichloroethene-d2	2.565	65	14816	5.442	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	108.800%	
20) 2-Butanone-d5	4.662	46	88576	50.367	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	100.740%	
24) Chloroform-d	5.064	84	68382	5.231	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	104.600%	
26) 1,2-Dichloroethane-d4	5.707	65	35561	5.398	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	108.000%	
32) Benzene-d6	5.723	84	134470	4.882	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	97.600%	
36) 1,2-Dichloropropane-d6	6.694	67	46463	5.149	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	103.000%	
41) Toluene-d8	7.896	98	106631	4.356	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	87.200%	
43) trans-1,3-Dichloroprop...	8.182	79	14713	4.653	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	93.000%	
46) 2-Hexanone-d5	8.642	63	55792	47.128	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	94.260%	
56) 1,1,2,2-Tetrachloroeth...	10.758	84	40288	5.314	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	106.200%	
66) 1,2-Dichlorobenzene-d4	12.195	152	35092	5.249	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	105.000%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072822\
Data File : VU050002.D
Acq On : 28 Jul 2022 22:32
Operator : SY/MD
Sample : VU0728WBL02
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK102

Quant Time: Jul 29 05:05:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR072722WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Jul 29 05:04:48 2022
Response via : Initial Calibration

