

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072822\
 Data File : VU049981.D
 Acq On : 28 Jul 2022 12:36
 Operator : SY/MD
 Sample : VU0728WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK101

Quant Time: Jul 29 02:35:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR072722WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Jul 29 02:27:55 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	93399	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	93010	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	35285	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	40998	5.113	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 102.200%		
7) Chloroethane-d5	1.913	69	30874	5.360	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	= 107.200%		
11) 1,1-Dichloroethene-d2	2.562	65	14723	5.327	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 106.600%		
20) 2-Butanone-d5	4.642	46	76311	42.748	ug/L	-0.02
Spiked Amount	50.000	Range 40 - 130	Recovery	= 85.500%		
24) Chloroform-d	5.064	84	67623	5.096	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 102.000%		
26) 1,2-Dichloroethane-d4	5.703	65	34822	5.208	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 104.200%		
32) Benzene-d6	5.729	84	135733	4.939	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 98.800%		
36) 1,2-Dichloropropane-d6	6.690	67	45826	5.090	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 101.800%		
41) Toluene-d8	7.899	98	110732	4.533	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 90.600%		
43) trans-1,3-Dichloroprop...	8.182	79	15698	4.975	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 99.400%		
46) 2-Hexanone-d5	8.639	63	56718	48.012	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 96.020%		
56) 1,1,2,2-Tetrachloroeth...	10.758	84	39381	5.206	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 104.200%		
66) 1,2-Dichlorobenzene-d4	12.195	152	35347	5.401	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 108.000%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072822\
Data File : VU049981.D
Acq On : 28 Jul 2022 12:36
Operator : SY/MD
Sample : VU0728WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK101

Quant Time: Jul 29 02:35:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR072722WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Jul 29 02:27:55 2022
Response via : Initial Calibration

