

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072822\
 Data File : VU050012.D
 Acq On : 29 Jul 2022 02:38
 Operator : SY/MD
 Sample : N3950-05
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0CF8

Quant Time: Jul 29 06:41:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR072722WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Jul 29 05:04:48 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	82343	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	84315	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	32023	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	29420	4.161	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	83.200%
7) Chloroethane-d5	1.912	69	24115	4.749	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	95.000%
11) 1,1-Dichloroethene-d2	2.565	65	10816	4.439	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	88.800%
20) 2-Butanone-d5	4.636	46	87277	55.456	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	110.920%
24) Chloroform-d	5.063	84	57169	4.886	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.800%
26) 1,2-Dichloroethane-d4	5.703	65	29991	5.087	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.800%
32) Benzene-d6	5.729	84	111978	4.494	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.800%
36) 1,2-Dichloropropane-d6	6.690	67	38383	4.702	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.000%
41) Toluene-d8	7.899	98	87584	3.955	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	79.200%
43) trans-1,3-Dichloroprop...	8.182	79	12133	4.241	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	84.800%
46) 2-Hexanone-d5	8.636	63	47023	43.910	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	87.820%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	34539	5.036	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	100.800%
66) 1,2-Dichlorobenzene-d4	12.195	152	30497	5.134	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	102.600%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072822\
Data File : VU050012.D
Acq On : 29 Jul 2022 02:38
Operator : SY/MD
Sample : N3950-05
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0CF8

Quant Time: Jul 29 06:41:41 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR072722WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Jul 29 05:04:48 2022
Response via : Initial Calibration

