

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072823\
 Data File : VU054909.D
 Acq On : 28 Jul 2023 13:53
 Operator : MD/SY
 Sample : 03587-09DL 5X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PT-THM-WSDL

Quant Time: Jul 31 11:18:24 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U072523DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Wed Jul 26 09:45:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Fluorobenzene	6.113	96	24017	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.634	95	7322	1.006	ug/l	0.00
Spiked Amount 1.000			Recovery	=	101.000%	
68) 1,2-Dichlorobenzene-d4	12.196	152	8143	0.990	ug/l	0.00
Spiked Amount 1.000			Recovery	=	99.000%	
Target Compounds						
27) Chloroform	5.084	83	48870	4.507	ug/l	98
44) Bromodichloromethane	7.103	83	50740	6.341	ug/l	99
55) Dibromochloromethane	8.808	129	11250	2.156	ug/l	98
67) Bromoform	10.290	173	19132	7.072	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072823\
Data File : VU054909.D
Acq On : 28 Jul 2023 13:53
Operator : MD/SY
Sample : 03587-09DL 5X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PT-THM-WSDL

Quant Time: Jul 31 11:18:24 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U072523DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Wed Jul 26 09:45:35 2023
Response via : Initial Calibration

