



Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072823\
Data File : VU054904.D
Acq On : 28 Jul 2023 09:32
Operator : MD/SY
Sample : BFB
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFB

Integration File: rteint.p

Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U072523DW.M
Title : METHOD 524.2 VOLATILES DRINKING WATER
Last Update : Wed Jul 26 09:45:35 2023



AutoFind: Scans 2905, 2906, 2907; Background Corrected with Scan 2893

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	24.0	4104	PASS
75	95	30	60	58.2	9959	PASS
95	95	100	100	100.0	17097	PASS
96	95	5	9	6.5	1111	PASS
173	174	0.00	2	1.6	257	PASS
174	95	50	100	97.0	16580	PASS
175	174	5	9	7.2	1187	PASS
176	174	95	101	97.9	16224	PASS
177	176	5	9	5.8	949	PASS