

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU072919\
 Data File : VU033474.D
 Acq On : 29 Jul 2019 17:41
 Operator : JC/SP
 Sample : VU0729WBL01
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK31

Quant Time: Jul 30 04:29:39 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SFAMULM072919W.M
 Quant Title : VOC Analysis
 QLast Update : Tue Jul 30 04:12:14 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.86	114	276938	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.07	117	283362	50.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.47	152	124720	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	125772	50.02	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	100.04%
7) Chloroethane-d5	1.67	69	109383	52.65	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	105.30%
11) 1,1-Dichloroethene-d2	2.26	63	158538	34.47	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	68.94%
21) 2-Butanone-d5	4.15	46	141884	98.46	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	98.46%
24) Chloroform-d	4.62	84	187474	49.82	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.64%
26) 1,2-Dichloroethane-d4	5.29	65	130354	53.69	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	107.38%
32) Benzene-d6	5.32	84	399545	52.53	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	105.06%
36) 1,2-Dichloropropane-d6	6.31	67	125057	53.00	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	106.00%
41) Toluene-d8	7.55	98	351896	49.21	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.42%
43) trans-1,3-Dichloropropene-	7.83	79	55301	47.96	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	95.92%
47) 2-Hexanone-d5	8.30	63	104668	106.79	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	106.79%
56) 1,1,2,2-Tetrachloroethane-	10.41	84	181501	50.12	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	100.24%
66) 1,2-Dichlorobenzene-d4	11.85	152	133343	55.64	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	111.28%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU072919\
Data File : VU033474.D
Acq On : 29 Jul 2019 17:41
Operator : JC/SP
Sample : VU0729WBL01
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK31

Quant Time: Jul 30 04:29:39 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SFAMULM072919W.M
Quant Title : VOC Analysis
QLast Update : Tue Jul 30 04:12:14 2019
Response via : Initial Calibration

