

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU072920\
 Data File : VU039701.D
 Acq On : 29 Jul 2020 13:36
 Operator : SY/MD
 Sample : I.BLK
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK37

Quant Time: Jul 30 01:53:56 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA U\METHOD\524U072920DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Thu Jul 30 01:45:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
-----	-----	-----	-----	-----	-----	-----
1) Fluorobenzene	6.13	96	22715	1.00	ug/l	0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.64	95	7429	0.84	ug/l	0.00
Spiked Amount 1.000			Recovery	=	84.00%	
68) 1,2-Dichlorobenzene-d4	12.19	152	7446	0.86	ug/l	0.00
Spiked Amount 1.000			Recovery	=	86.00%	
Target Compounds						Ovalue
45) 4-Methyl-2-Pentanone	7.80	43	3106	0.486	ug/l	98
54) 2-Hexanone	8.70	43	3065	0.670	ug/l	91
89) Naphthalene	14.08	128	6357	0.325	ug/l	99
-----	-----	-----	-----	-----	-----	-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU072920\
Data File : VU039701.D
Acq On : 29 Jul 2020 13:36
Operator : SY/MD
Sample : I.BLK
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK37

Quant Time: Jul 30 01:53:56 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA U\METHOD\524U072920DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Thu Jul 30 01:45:12 2020
Response via : Initial Calibration

