

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU072920\
 Data File : VU039705.D
 Acq On : 29 Jul 2020 15:56
 Operator : SY/MD
 Sample : L3319-11
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PT-EDBCP-WS

Quant Time: Jul 30 09:26:06 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA U\METHOD\524U072920DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Thu Jul 30 01:45:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Fluorobenzene	6.13	96	20806	1.00	ug/l	0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.63	95	6267	0.77	ug/l	0.00
Spiked Amount 1.000			Recovery	=	77.00%	
68) 1,2-Dichlorobenzene-d4	12.19	152	6562	0.82	ug/l	0.00
Spiked Amount 1.000			Recovery	=	82.00%	
Target Compounds						
56) 1,2-Dibromoethane	8.93	107	6338	1.311	ug/l	96
71) 1,2,3-Trichloropropane	10.83	75	6287	1.156	ug/l	88
86) 1,2-Dibromo-3-Chloropropan	12.99	75	1313	1.079	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU072920\
Data File : VU039705.D
Acq On : 29 Jul 2020 15:56
Operator : SY/MD
Sample : L3319-11
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PT-EDBCP-WS

Quant Time: Jul 30 09:26:06 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA U\METHOD\524U072920DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Thu Jul 30 01:45:12 2020
Response via : Initial Calibration

