

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU072920\
 Data File : VU039707.D
 Acq On : 29 Jul 2020 16:54
 Operator : SY/MD
 Sample : L3319-10
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PT-ADD-WS

Quant Time: Jul 30 09:27:03 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA U\METHOD\524U072920DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Thu Jul 30 01:45:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Fluorobenzene	6.13	96	21172	1.00	ug/l	0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.63	95	6511	0.79	ug/l	0.00
Spiked Amount 1.000			Recovery	=	79.00%	
68) 1,2-Dichlorobenzene-d4	12.19	152	7353	0.91	ug/l	0.00
Spiked Amount 1.000			Recovery	=	91.00%	
Target Compounds						
					Ovalue	
8) 1,1,2-Trichloro-1,2,2-trif	2.59	101	146819	22.964	ug/l	99
24) tert-Butyl Alcohol	3.20	59	34112	42.883	ug/l	98
25) Methyl tert-Butyl Ether	3.39	73	318572	16.597	ug/l	99
31) Isopropyl Ether	4.02	45	295478	11.031	ug/l	99
32) Ethyl-t-butyl ether	4.53	59	274514	12.043	ug/l	98
33) Tert-Amyl methyl ether	5.97	73	361673	22.008	ug/l	100
73) n-propylbenzene	10.91	120	398704	48.978	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU072920\
Data File : VU039707.D
Acq On : 29 Jul 2020 16:54
Operator : SY/MD
Sample : L3319-10
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PT-ADD-WS

Quant Time: Jul 30 09:27:03 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA U\METHOD\524U072920DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Thu Jul 30 01:45:12 2020
Response via : Initial Calibration

