

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU072920\
 Data File : VU039710.D
 Acq On : 29 Jul 2020 18:24
 Operator : SY/MD
 Sample : VIBLK
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK

Quant Time: Jul 30 02:37:26 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA U\METHOD\524U072920DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Thu Jul 30 01:45:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Fluorobenzene	6.12	96	21240	1.00	ug/l	0.00

System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.63	95	6723	0.81	ug/l	0.00
Spiked Amount 1.000			Recovery	=	81.00%	
68) 1,2-Dichlorobenzene-d4	12.19	152	6625	0.82	ug/l	0.00
Spiked Amount 1.000			Recovery	=	82.00%	

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU072920\
Data File : VU039710.D
Acq On : 29 Jul 2020 18:24
Operator : SY/MD
Sample : VIBLK
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK

Quant Time: Jul 30 02:37:26 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA U\METHOD\524U072920DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Thu Jul 30 01:45:12 2020
Response via : Initial Calibration

