

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072922\
 Data File : VU050026.D
 Acq On : 29 Jul 2022 10:43
 Operator : SY/MD
 Sample : VU0729WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK103

Quant Time: Jul 30 01:41:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR072722WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Jul 30 01:40:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	86913	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	86850	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	33502	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	32300	4.328	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	86.600%	
7) Chloroethane-d5	1.912	69	27270	5.088	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	101.800%	
11) 1,1-Dichloroethene-d2	2.565	65	11499	4.471	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	89.400%	
20) 2-Butanone-d5	4.636	46	80870	48.683	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	97.360%	
24) Chloroform-d	5.063	84	61530	4.983	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	99.600%	
26) 1,2-Dichloroethane-d4	5.703	65	32935	5.293	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	105.800%	
32) Benzene-d6	5.726	84	119792	4.668	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	93.400%	
36) 1,2-Dichloropropane-d6	6.694	67	41728	4.963	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	99.200%	
41) Toluene-d8	7.899	98	94173	4.128	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	82.600%	
43) trans-1,3-Dichloroprop...	8.179	79	14344	4.868	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	97.400%	
46) 2-Hexanone-d5	8.639	63	53012	48.058	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	96.120%	
56) 1,1,2,2-Tetrachloroeth...	10.758	84	36667	5.191	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	103.800%	
66) 1,2-Dichlorobenzene-d4	12.195	152	30940	4.979	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	99.600%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072922\
Data File : VU050026.D
Acq On : 29 Jul 2022 10:43
Operator : SY/MD
Sample : VU0729WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK103

Quant Time: Jul 30 01:41:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR072722WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jul 30 01:40:38 2022
Response via : Initial Calibration

